



Ref. Certif. No.
SE-105641

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Medical Power Supply

Name and address of the applicant

Applied Power (Shanghai) Electronics Co., Ltd.
No.400 Fangchun Rd., free trade zone
Shanghai 201203, China

Name and address of the manufacturer

Same as applicant

Name and address of the factory

Note: When more than one factory, please report on page 2

Chenzhou Frecom Electronics Co., Ltd.
No.7 Bldg, Chasing Dream Park, Jiangbei IZ, Chenzhou Zixing,
Hunan 423400, China

Ratings and principal characteristics

Input: 100-240V~, 50-60Hz, 2.0A
Output: 12V@5A, 15V@4A, 19V@3.2A, 24V@2.5A, 28V@2.1A,
48V@1.3A

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

MA60*****

Additional information (if necessary may also be reported on page 2)

See page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1+A2

As shown in the Test Report Ref. No. which forms part of this Certificate

210402145SHA-001

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Torshamnsgatan 43
Box 1103
SE-164 22 Kista, Sweden

intertek

Signature:

Gary Hu

Date: 10 August, 2021



Ref. Certif. No.

SE-105641

Additional information

Explanation of model MA60*****:

The first symbol “*” represents inlet options: A for C14, B for C8, C for C6, D for C18.

The second symbol “*” represents output connector options: code A-Z for different type of connectors.

The third and fourth symbol “**” represents output voltages: 12 for 12V, 15 for 15V, 19 for 19V, 24 for 24V, 28 for 28V, 48 for 48V, +/-20% tolerance.

The last 2 symbols “**” can be AA, AB, AC, AE, AF, BA, BB, BC, BE, BF, CA, CB, CC, CE, CF, represents internal control.

The national differences for Canada, USA have been checked.

Following applicable requirement have not been evaluated by Intertek:

Biocompatibility according to ISO 10993

EMC according to IEC 60601-1-2

Date: 10 August, 2021

Signature: